
Midmark M11 - v1.0.5
Total Cycles: 172

Name: RITTER2
Sterilizer ID: V2313228
07/07/2021 09:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.5 F 0.0
2:00 194.5 F 3.5
4:00 202.7 F 9.0
6:00 219.4 F 8.1
8:00 229.9 F 10.6
10:00 235.0 F 8.7
12:00 239.6 F 10.1
14:00 249.4 F 15.1
16:00 261.7 F 22.2
18:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 271.8 F 29.0
2:00 272.0 F 29.3
3:00 272.1 F 29.6
4:00 271.3 F 28.8

Min 271.1 F 28.8
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.4 F 0.0
5:00 236.7 F 0.0
10:00 227.1 F 0.0
15:00 218.7 F 0.0
20:00 230.0 F 0.0
25:00 229.4 F 0.0
30:00 231.0 F 0.0

Drying Complete

FILLING: 2:14
HEATING: 18:20
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:41
CYCLE COMPLETE

07/07/2021 10:14 AM
ACCEPTED: X
REJECTED:
Operator: U01