
Midmark M11 - v1.0.0
Total Cycles: 3261

Name: RITTER 1
Sterilizer ID: V2178198
05/19/2023 09:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.5 F 0.0
2:00 163.6 F 0.0
4:00 165.9 F 2.0
6:00 195.7 F 6.2
8:00 212.5 F 11.4
10:00 223.4 F 6.1
12:00 234.9 F 11.3
14:00 234.4 F 8.3
16:00 241.9 F 10.9
18:00 245.9 F 13.1
20:00 258.3 F 19.7
22:00 269.2 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.7 F 29.0
2:00 271.9 F 28.3
3:00 271.8 F 28.6
4:00 272.7 F 29.0

Min 271.4 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 224.9 F 0.0
10:00 237.2 F 0.0
15:00 225.9 F 0.0
20:00 226.3 F 0.0
25:00 225.1 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 23:06
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:01:15
CYCLE COMPLETE

05/19/2023 10:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____