

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1281

Name: RITTER2  
Sterilizer ID: V2313228  
07/11/2022 12:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.1 F 0.0  
2:00 144.2 F 1.6  
4:00 170.1 F 4.3  
6:00 188.3 F 8.2  
8:00 219.8 F 12.5  
10:00 214.2 F 6.1  
12:00 229.9 F 10.7  
14:00 231.1 F 7.8  
16:00 240.5 F 12.3  
18:00 238.8 F 10.2  
20:00 250.6 F 16.2  
22:00 261.4 F 22.4  
24:00 271.2 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.0 F 30.1  
2:00 272.5 F 29.8  
3:00 272.8 F 30.3  
4:00 272.0 F 30.2

Min 270.7 F 29.0  
Max 273.2 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 195.8 F 0.0  
10:00 208.4 F 0.0  
15:00 211.5 F 0.0  
20:00 206.9 F 0.0  
25:00 206.2 F 0.0  
30:00 205.3 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 24:37  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:02:48  
CYCLE COMPLETE

07/11/2022 01:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_