
Midmark M11 - v1.0.5
Total Cycles: 3531

Name: RITTER2
Sterilizer ID: V2313228
04/16/2024 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.0
2:00 83.3 F 0.0
4:00 121.2 F 0.0
6:00 117.4 F 0.2
8:00 148.3 F 2.4
10:00 172.6 F 5.2
12:00 197.7 F 8.7
14:00 216.9 F 12.5
16:00 220.3 F 5.4
18:00 231.5 F 9.2
20:00 240.4 F 13.5
22:00 237.6 F 10.2
24:00 236.9 F 8.9
26:00 247.7 F 14.2
28:00 258.8 F 20.4
30:00 269.1 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.5 F 29.0
2:00 272.5 F 29.5
3:00 271.8 F 28.6
4:00 272.6 F 29.3

Min 271.6 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.7 F 0.0
5:00 222.5 F 0.0
10:00 199.3 F 0.0
15:00 185.9 F 0.0
20:00 179.3 F 0.0
25:00 175.8 F 0.0
30:00 174.5 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:05
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:09:11
CYCLE COMPLETE

04/16/2024 10:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____