
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001594
14:51 15/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:16
STERILIZING 3:17
136.0 C 220kPa 3:17
136.0 C 223kPa 3:47
136.3 C 224kPa 4:17
136.0 C 222kPa 4:47
136.3 C 225kPa 5:17
136.0 C 222kPa 5:47
136.1 C 224kPa 6:17
136.4 C 224kPa 6:47
136.4 C 224kPa 6:47
VENTING 6:47
AIR DRYING 7:42
CYCLE COMPLETE 7:51

Digital Signature #
848283958CE577E1
