
Midmark M11 - v1.0.5
Total Cycles: 2609

Name: RITTER2
Sterilizer ID: V2313228
07/05/2023 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.3 F 0.0
2:00 196.2 F 4.0
4:00 213.9 F 9.9
6:00 224.5 F 6.9
8:00 237.1 F 12.1
10:00 237.5 F 10.2
12:00 236.8 F 9.1
14:00 251.2 F 16.2
16:00 263.2 F 23.3
18:00 272.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.4
2:00 271.5 F 28.8
3:00 272.9 F 29.6
4:00 271.9 F 28.9

Min 271.4 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 220.1 F 0.0
10:00 236.9 F 0.0
15:00 236.4 F 0.0
20:00 227.4 F 0.0
25:00 223.7 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 2:31
HEATING: 18:08
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

07/05/2023 11:55 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____