

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2054

Name: RITTER2  
Sterilizer ID: V2313228  
02/15/2023 10:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.7 F 0.0  
2:00 180.7 F 1.7  
4:00 181.1 F 5.4  
6:00 204.5 F 10.5  
8:00 218.7 F 5.8  
10:00 234.0 F 11.7  
12:00 234.8 F 9.5  
14:00 236.6 F 9.1  
16:00 249.1 F 15.5  
18:00 261.5 F 22.5  
20:00 272.7 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.5 F 29.6  
2:00 271.6 F 29.5  
3:00 272.4 F 29.6  
4:00 271.4 F 29.4

Min 271.3 F 29.0  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.3 F 0.0  
5:00 214.8 F 0.0  
10:00 219.4 F 0.0  
15:00 225.0 F 0.0  
20:00 227.6 F 0.0  
25:00 229.8 F 0.0  
30:00 227.0 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 20:24  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:53  
CYCLE COMPLETE

02/15/2023 11:04 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_