
Midmark M11 - v1.0.0
Total Cycles: 5263

Name: RITTER 1
Sterilizer ID: V2178198
01/22/2025 07:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.0 F 0.0
2:00 167.3 F 3.3
4:00 182.4 F 7.6
6:00 215.8 F 12.7
8:00 223.2 F 5.9
10:00 236.1 F 10.6
12:00 232.5 F 7.3
14:00 244.4 F 12.9
16:00 242.8 F 11.4
18:00 255.5 F 17.9
20:00 267.0 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.5 F 28.8
2:00 271.6 F 28.0
3:00 272.1 F 28.8
4:00 272.2 F 28.6

Min 271.2 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 198.8 F 0.0
10:00 194.0 F 0.0
15:00 200.5 F 0.0
20:00 226.8 F 0.0
25:00 222.9 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:01:27
CYCLE COMPLETE

01/22/2025 08:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____