

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1817

Name: RITTER 1  
Sterilizer ID: V2178198  
04/13/2022 08:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 211.2 F 0.0  
2:00 193.6 F 2.9  
4:00 194.6 F 7.3  
6:00 212.7 F 12.7  
8:00 224.2 F 7.1  
10:00 236.8 F 12.3  
12:00 236.8 F 9.5  
14:00 237.6 F 8.8  
16:00 248.7 F 14.5  
18:00 260.9 F 21.4  
20:00 272.2 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.3 F 28.5  
2:00 272.3 F 28.9  
3:00 271.7 F 28.2  
4:00 272.5 F 28.9  
  
Min 271.4 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.9 F 0.0  
5:00 227.6 F 0.0  
10:00 216.3 F 0.0  
15:00 215.2 F 0.0  
20:00 229.6 F 0.0  
25:00 241.8 F 0.0  
30:00 224.9 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 20:27  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:55  
CYCLE COMPLETE

04/13/2022 09:42 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_