

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1015

Name: RITTER2  
Sterilizer ID: V2313228  
04/21/2022 06:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.8 F 0.0  
2:00 74.0 F 0.0  
4:00 76.1 F 0.0  
6:00 108.2 F 0.0  
8:00 142.1 F 1.3  
10:00 166.6 F 4.0  
12:00 183.2 F 6.8  
14:00 203.4 F 10.7  
16:00 219.8 F 5.7  
18:00 228.8 F 8.2  
20:00 239.1 F 12.5  
22:00 236.2 F 9.4  
24:00 239.6 F 10.1  
26:00 246.7 F 13.9  
28:00 258.3 F 20.1  
30:00 269.0 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.7 F 29.3  
2:00 271.2 F 28.6  
3:00 272.7 F 29.4  
4:00 271.5 F 28.6

Min 271.1 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.8 F 0.0  
5:00 213.5 F 0.0  
10:00 209.6 F 0.0  
15:00 199.1 F 0.0  
20:00 205.3 F 0.0  
25:00 215.9 F 0.0  
30:00 208.7 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 31:07  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:09:15  
CYCLE COMPLETE

04/21/2022 08:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_