
Midmark M11 - v1.0.0
Total Cycles: 4419

Name: RITTER 1
Sterilizer ID: V2178198
04/01/2024 09:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.0 F 0.0
2:00 75.1 F 0.2
4:00 85.9 F 1.2
6:00 126.4 F 3.6
8:00 152.1 F 7.0
10:00 175.0 F 11.1
12:00 194.2 F 2.2
14:00 217.9 F 5.4
16:00 229.1 F 8.9
18:00 238.4 F 13.0
20:00 233.6 F 8.6
22:00 243.5 F 12.0
24:00 243.1 F 11.8
26:00 255.7 F 18.3
28:00 266.9 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.2 F 28.3
2:00 272.6 F 28.9
3:00 271.5 F 28.0
4:00 272.4 F 28.5

Min 271.3 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 219.8 F 0.0
10:00 226.2 F 0.0
15:00 228.5 F 0.0
20:00 231.6 F 0.0
25:00 228.1 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:29
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:07:26
CYCLE COMPLETE

04/01/2024 10:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____