

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007133  
10:54 28/03/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:53  
PRESSURIZING 1:56  
STERILIZING 2:54  
136.0 C 220kPa 2:54  
136.2 C 217kPa 3:24  
136.2 C 221kPa 3:54  
136.4 C 222kPa 4:24  
136.2 C 222kPa 4:54  
136.2 C 223kPa 5:24  
136.4 C 221kPa 5:54  
136.4 C 219kPa 6:24  
136.4 C 219kPa 6:24  
VENTING 6:24  
AIR DRYING 7:14  
CYCLE COMPLETE 7:21  
-----

Digital Signature #  
D5B65E976F54780B  
-----