

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002284  
13:20 18/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:20  
PRESSURIZING 2:25  
STERILIZING 3:22  
136.4 C 224kPa 3:22  
136.2 C 224kPa 3:52  
136.4 C 224kPa 4:22  
136.4 C 226kPa 4:52  
136.4 C 225kPa 5:22  
136.4 C 227kPa 5:52  
136.2 C 225kPa 6:22  
136.3 C 225kPa 6:52  
136.3 C 225kPa 6:52  
VENTING 6:52  
AIR DRYING 7:49  
CYCLE COMPLETE 8:04  
-----

Digital Signature #  
847CA1D12FD41F86  
-----