
Midmark M11 - v1.0.5
Total Cycles: 538

Name: RITTER2
Sterilizer ID: V2313228
11/10/2021 03:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.0 F 0.0
2:00 202.2 F 4.1
4:00 204.6 F 9.8
6:00 222.1 F 6.9
8:00 232.9 F 11.3
10:00 234.3 F 9.3
12:00 237.9 F 9.6
14:00 248.4 F 15.3
16:00 260.5 F 22.1
18:00 271.3 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.5 F 29.7
2:00 272.1 F 30.1
3:00 272.0 F 29.4
4:00 272.4 F 30.1

Min 271.5 F 29.3
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 224.3 F 0.0
10:00 226.3 F 0.0
15:00 222.8 F 0.0
20:00 222.0 F 0.0
25:00 221.4 F 0.0
30:00 219.9 F 0.0
Drying Complete

FILLING: 2:54
HEATING: 18:35
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

11/11/2021 07:26 AM
ACCEPTED: X
REJECTED:
Operator: U02