

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 008036  
12:33 07/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:22  
PRESSURIZING 2:25  
STERILIZING 3:30  
136.5 C 222kPa 3:30  
136.1 C 221kPa 4:00  
136.4 C 224kPa 4:30  
136.2 C 222kPa 5:00  
136.4 C 223kPa 5:30  
136.4 C 222kPa 6:00  
136.2 C 218kPa 6:30  
136.4 C 222kPa 7:00  
136.4 C 222kPa 7:00  
VENTING 7:00  
AIR DRYING 7:49  
CYCLE COMPLETE 37:49

-----  
Digital Signature #  
DEE3816F8CD663D9  
-----