

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3392

Name: RITTER2  
Sterilizer ID: V2313228  
02/16/2024 02:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.6 F 0.0  
2:00 122.3 F 0.3  
4:00 137.8 F 2.1  
6:00 168.7 F 5.2  
8:00 183.6 F 8.9  
10:00 201.5 F 12.9  
12:00 217.4 F 6.4  
14:00 227.4 F 10.7  
16:00 230.7 F 7.9  
18:00 240.6 F 12.3  
20:00 239.7 F 10.8  
22:00 251.7 F 16.7  
24:00 262.5 F 23.0  
26:00 272.4 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 273.0 F 30.0  
2:00 272.7 F 30.1  
3:00 271.9 F 29.9  
4:00 271.7 F 29.5

Min 270.9 F 29.0  
Max 273.1 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.7 F 0.0  
5:00 195.8 F 0.0  
10:00 199.5 F 0.0  
15:00 199.0 F 0.0  
20:00 196.5 F 0.0  
25:00 194.2 F 0.0  
30:00 191.9 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 26:21  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:04:55  
CYCLE COMPLETE

02/16/2024 03:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_