

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 005051  
10:25 02/02/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:36  
STERILIZING 3:25  
136.0 C 220kPa 3:25  
136.3 C 218kPa 3:55  
136.4 C 222kPa 4:25  
136.4 C 217kPa 4:55  
136.4 C 223kPa 5:25  
136.4 C 218kPa 5:55  
136.3 C 224kPa 6:25  
136.2 C 222kPa 6:55  
136.2 C 222kPa 6:55  
VENTING 6:55  
AIR DRYING 7:47  
CYCLE COMPLETE 12:25

-----  
Digital Signature #  
D7E370CAD270B923  
-----