
Midmark M11 - v1.0.5
Total Cycles: 4384

Name: RITTER2
Sterilizer ID: V2313228
02/26/2025 03:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.1 F 0.0
2:00 160.6 F 2.3
4:00 175.5 F 6.2
6:00 200.4 F 10.3
8:00 223.8 F 7.3
10:00 226.6 F 8.9
12:00 237.4 F 12.4
14:00 237.9 F 11.2
16:00 237.8 F 9.8
18:00 250.1 F 16.0
20:00 261.1 F 22.4
22:00 271.2 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.2 F 30.2
2:00 272.8 F 30.0
3:00 272.3 F 30.0
4:00 272.8 F 30.3

Min 271.0 F 29.3
Max 273.5 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.6 F 0.0
5:00 206.7 F 0.0
10:00 208.4 F 0.0
15:00 207.6 F 0.0
20:00 211.1 F 0.0
25:00 211.9 F 0.0
30:00 209.6 F 0.0
Drying Complete

FILLING: 2:41
HEATING: 22:49
STERILIZING: 4:00
VENTING: 4:04
DRYING: 30:00
TOTAL CYCLE: 01:03:34
CYCLE COMPLETE

02/26/2025 04:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____