

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2024

Name: RITTER2  
Sterilizer ID: V2313228  
02/08/2023 10:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.3 F 0.0  
2:00 206.9 F 2.7  
4:00 206.6 F 8.1  
6:00 221.6 F 10.5  
8:00 232.1 F 10.0  
10:00 234.5 F 8.3  
12:00 240.7 F 10.7  
14:00 247.9 F 14.6  
16:00 260.8 F 21.9  
18:00 272.3 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 272.7 F 29.5  
2:00 271.9 F 29.4  
3:00 272.6 F 29.5  
4:00 271.8 F 29.2  
  
Min 271.7 F 28.9  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
5:00 214.7 F 0.0  
10:00 240.9 F 0.0  
15:00 236.2 F 0.0  
20:00 228.0 F 0.0  
25:00 227.3 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 2:05  
HEATING: 18:28  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:57:41  
CYCLE COMPLETE

02/08/2023 11:50 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_