
Midmark M11 - v1.0.5
Total Cycles: 4191

Name: RITTER2
Sterilizer ID: V2313228
12/05/2024 02:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.8 F 0.0
2:00 173.9 F 1.9
4:00 201.3 F 6.1
6:00 221.1 F 10.7
8:00 226.8 F 6.5
10:00 235.2 F 11.2
12:00 233.9 F 9.0
14:00 239.9 F 10.5
16:00 246.8 F 14.3
18:00 258.3 F 20.6
20:00 268.6 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.3 F 30.0
2:00 271.9 F 29.4
3:00 271.1 F 29.3
4:00 272.1 F 30.0

Min 271.1 F 29.0
Max 272.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 218.5 F 0.0
10:00 199.6 F 0.0
15:00 195.9 F 0.0
20:00 193.0 F 0.0
25:00 204.9 F 0.0
30:00 215.8 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 21:15
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:01:04
CYCLE COMPLETE

12/05/2024 03:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____