
Midmark M11 - v1.0.5
Total Cycles: 4369

Name: RITTER2
Sterilizer ID: V2313228
02/21/2025 11:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.6 F 0.0
2:00 115.2 F 0.9
4:00 129.6 F 3.3
6:00 155.0 F 6.7
8:00 181.4 F 10.6
10:00 200.3 F 5.6
12:00 214.6 F 6.3
14:00 227.9 F 10.6
16:00 229.4 F 7.9
18:00 241.1 F 12.3
20:00 238.9 F 10.5
22:00 251.1 F 16.4
24:00 261.8 F 22.6
26:00 271.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.1 F 29.9
2:00 272.4 F 29.9
3:00 271.6 F 29.5
4:00 272.3 F 29.8

Min 271.0 F 29.0
Max 273.4 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 232.8 F 0.0
10:00 223.2 F 0.0
15:00 215.2 F 0.0
20:00 221.8 F 0.0
25:00 230.0 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:44
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:05:24
CYCLE COMPLETE

02/21/2025 12:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____