
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 007686
13:45 16/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:46
STERILIZING 4:59
136.5 C 220kPa 4:59
136.3 C 224kPa 5:29
136.3 C 219kPa 5:59
136.2 C 223kPa 6:29
136.2 C 223kPa 6:59
136.0 C 221kPa 7:29
136.1 C 219kPa 7:59
136.4 C 221kPa 8:29
136.4 C 221kPa 8:29
VENTING 8:29
AIR DRYING 9:23
CYCLE COMPLETE 9:31

Digital Signature #
BB3DDFCEB9657643
