
Midmark M11 - v1.0.5
Total Cycles: 1660

Name: RITTER2
Sterilizer ID: V2313228
10/26/2022 11:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.8 F 0.0
2:00 179.9 F 3.6
4:00 186.5 F 8.0
6:00 209.2 F 13.3
8:00 225.7 F 8.0
10:00 235.6 F 12.4
12:00 238.8 F 11.4
14:00 238.6 F 10.4
16:00 252.5 F 17.4
18:00 264.4 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.6 F 29.9
2:00 271.6 F 29.2
3:00 272.0 F 30.0
4:00 272.3 F 29.7

Min 271.2 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 221.7 F 0.0
10:00 234.8 F 0.0
15:00 237.5 F 0.0
20:00 226.2 F 0.0
25:00 221.1 F 0.0
30:00 219.3 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 19:59
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

10/26/2022 12:32 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____