
Midmark M11 - v1.0.5
Total Cycles: 1691

Name: RITTER2
Sterilizer ID: V2313228
11/07/2022 04:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.4 F 0.0
2:00 165.9 F 0.3
4:00 186.6 F 1.9
6:00 207.1 F 6.5
8:00 224.0 F 11.7
10:00 228.9 F 8.1
12:00 241.3 F 12.3
14:00 242.4 F 12.2
16:00 241.8 F 11.4
18:00 255.0 F 18.3
20:00 266.8 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 273.0 F 29.6
2:00 271.8 F 28.8
3:00 272.0 F 29.4
4:00 272.4 F 29.4

Min 271.3 F 28.6
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.3 F 0.0
5:00 226.2 F 0.0
10:00 233.0 F 0.0
15:00 228.9 F 0.0
20:00 230.5 F 0.0
25:00 226.8 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:33
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:60
CYCLE COMPLETE

11/07/2022 05:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____