
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001000
14:18 14/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:36
STERILIZING 3:46
136.4 C 233kPa 3:46
136.2 C 232kPa 4:16
136.2 C 230kPa 4:46
136.4 C 233kPa 5:16
136.4 C 233kPa 5:46
136.4 C 233kPa 6:16
136.4 C 233kPa 6:46
136.3 C 231kPa 7:16
136.3 C 231kPa 7:16
VENTING 7:16
AIR DRYING 8:07
CYCLE COMPLETE 8:22

Digital Signature #
205A9F823C8BA9A4
