
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001277
12:40 25/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:03
STERILIZING 4:01
136.4 C 223kPa 4:01
136.3 C 224kPa 4:31
136.3 C 223kPa 5:01
136.1 C 218kPa 5:31
136.4 C 224kPa 6:01
136.4 C 224kPa 6:31
136.4 C 224kPa 7:01
136.4 C 225kPa 7:31
136.4 C 225kPa 7:31
VENTING 7:31
AIR DRYING 8:23
CYCLE COMPLETE 8:28

Digital Signature #
10E3A45300997B96
