

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002509  
10:38 11/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:07  
PRESSURIZING 3:09  
STERILIZING 4:16  
136.3 C 222kPa 4:16  
136.1 C 223kPa 4:46  
136.4 C 225kPa 5:16  
136.4 C 225kPa 5:46  
136.4 C 223kPa 6:16  
136.2 C 224kPa 6:46  
136.4 C 225kPa 7:16  
136.3 C 225kPa 7:46  
136.3 C 225kPa 7:46  
VENTING 7:46  
AIR DRYING 8:41  
CYCLE COMPLETE 8:45

-----  
Digital Signature #  
CC64B3A05280030B  
-----