

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1582

Name: RITTER 1  
Sterilizer ID: V2178198  
02/01/2022 01:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.5 F 0.0  
2:00 179.0 F 3.3  
4:00 186.2 F 8.0  
6:00 203.5 F 12.9  
8:00 223.3 F 7.0  
10:00 236.5 F 12.2  
12:00 235.3 F 9.1  
14:00 239.9 F 9.6  
16:00 247.6 F 13.8  
18:00 259.7 F 20.5  
20:00 270.6 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.6 F 28.8  
2:00 271.7 F 28.1  
3:00 271.7 F 28.5  
4:00 272.6 F 28.7  
  
Min 271.1 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
5:00 244.4 F 0.0  
10:00 228.7 F 0.0  
15:00 227.4 F 0.0  
20:00 229.4 F 0.0  
25:00 230.3 F 0.0  
30:00 230.5 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 20:51  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:32  
CYCLE COMPLETE

02/01/2022 02:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_