
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001498
16:28 04/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 2:02
STERILIZING 2:56
136.0 C 220kPa 2:56
136.4 C 225kPa 3:26
136.3 C 225kPa 3:56
136.4 C 226kPa 4:26
136.3 C 225kPa 4:56
136.3 C 225kPa 5:26
136.1 C 223kPa 5:56
136.4 C 225kPa 6:26
136.4 C 225kPa 6:26
VENTING 6:26
AIR DRYING 7:18
CYCLE COMPLETE 37:18

Digital Signature #
CC2CE7D17934549F
