

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000392  
10:50 25/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 0:57  
PRESSURIZING 1:07  
STERILIZING 1:55  
136.0 C 232kPa 1:55  
136.2 C 232kPa 2:25  
136.0 C 230kPa 2:55  
136.3 C 233kPa 3:25  
136.4 C 234kPa 3:55  
136.4 C 231kPa 4:25  
136.4 C 232kPa 4:55  
136.4 C 232kPa 5:25  
136.4 C 232kPa 5:25  
VENTING 5:25  
AIR DRYING 6:13  
CYCLE COMPLETE 6:56  
-----

Digital Signature #  
5E28A80CC6AC0CC4  
-----