
Midmark M11 - v1.0.5
Total Cycles: 373

Name: RITTER2
Sterilizer ID: V2313228
09/16/2021 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.1 F 0.0
2:00 183.1 F 2.6
4:00 183.3 F 6.6
6:00 198.8 F 11.3
8:00 216.9 F 6.2
10:00 231.2 F 11.1
12:00 232.4 F 8.3
14:00 241.9 F 11.5
16:00 244.5 F 13.2
18:00 256.3 F 19.5
20:00 266.8 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.5
1:00 272.7 F 30.2
2:00 271.5 F 29.4
3:00 271.7 F 30.2
4:00 273.6 F 30.6

Min 270.9 F 29.4
Max 273.6 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.5 F 0.0
5:00 234.6 F 0.0
10:00 222.1 F 0.0
15:00 216.0 F 0.0
20:00 214.0 F 0.0
25:00 217.1 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 21:35
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:00:25
CYCLE COMPLETE

09/16/2021 01:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____