

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1829

Name: RITTER 1  
Sterilizer ID: V2178198  
04/14/2022 04:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.3 F 0.0  
2:00 196.3 F 4.5  
4:00 212.0 F 10.2  
6:00 225.7 F 5.7  
8:00 235.9 F 11.9  
10:00 236.3 F 9.6  
12:00 238.5 F 8.7  
14:00 250.1 F 15.5  
16:00 262.7 F 22.8  
18:00 271.9 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.6 F 29.2  
2:00 271.5 F 28.5  
3:00 272.6 F 29.4  
4:00 272.3 F 28.8  
  
Min 271.3 F 28.5  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.7 F 0.0  
5:00 214.2 F 0.0  
10:00 230.4 F 0.0  
15:00 233.5 F 0.0  
20:00 227.4 F 0.0  
25:00 224.6 F 0.0  
30:00 224.3 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 18:09  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:18  
CYCLE COMPLETE

04/14/2022 05:03 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_