
Midmark M11 - v1.0.5
Total Cycles: 2248

Name: RITTER2
Sterilizer ID: V2313228
04/13/2023 01:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.3 F 0.0
2:00 179.5 F 3.4
4:00 191.5 F 8.1
6:00 212.4 F 13.2
8:00 224.8 F 8.6
10:00 235.6 F 11.1
12:00 240.6 F 12.2
14:00 240.6 F 11.3
16:00 254.2 F 18.2
18:00 265.9 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 271.8 F 29.6
2:00 272.3 F 29.6
3:00 271.5 F 29.0
4:00 271.7 F 29.7

Min 271.2 F 28.9
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 218.4 F 0.0
10:00 230.1 F 0.0
15:00 225.2 F 0.0
20:00 222.0 F 0.0
25:00 220.3 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 19:32
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:36
CYCLE COMPLETE

04/13/2023 02:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____