
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000938
13:38 17/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:18
PRESSURIZING 4:19
STERILIZING 5:36
136.0 C 227kPa 5:36
136.2 C 230kPa 6:06
136.4 C 231kPa 6:36
136.4 C 233kPa 7:06
136.0 C 229kPa 7:36
136.2 C 231kPa 8:06
136.3 C 232kPa 8:36
136.4 C 232kPa 9:06
136.4 C 232kPa 9:06
VENTING 9:06
AIR DRYING 10:05
CYCLE COMPLETE 25:05

Digital Signature #
42642E1E6AD26164
