
Midmark M11 - v1.0.5
Total Cycles: 3981

Name: RITTER2
Sterilizer ID: V2313228
10/01/2024 08:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.5 F 0.0
2:00 70.8 F 0.0
4:00 73.4 F 0.0
6:00 110.0 F 0.1
8:00 133.2 F 0.3
10:00 156.3 F 0.9
12:00 180.1 F 3.4
14:00 207.2 F 5.9
16:00 211.2 F 9.2
18:00 223.0 F 12.8
20:00 224.8 F 7.3
22:00 234.4 F 11.6
24:00 233.2 F 8.4
26:00 244.1 F 12.6
28:00 243.4 F 12.2
30:00 254.3 F 17.8
32:00 264.1 F 23.7
34:00 272.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 272.4 F 29.1
2:00 272.7 F 29.4
3:00 272.8 F 29.6
4:00 272.3 F 29.7

Min 271.9 F 29.1
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.9 F 0.0
5:00 211.0 F 0.0
10:00 216.3 F 0.0
15:00 183.7 F 0.0
20:00 182.3 F 0.0
25:00 199.2 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 34:58
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:13:23
CYCLE COMPLETE

10/01/2024 10:10 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____