
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 006201
14:11 29/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:29
PRESSURIZING 3:29
STERILIZING 4:29
136.4 C 221kPa 4:29
136.4 C 224kPa 4:59
136.2 C 218kPa 5:29
135.9 C 220kPa 5:59
136.3 C 224kPa 6:29
136.3 C 223kPa 6:59
136.2 C 222kPa 7:29
136.1 C 221kPa 7:59
136.1 C 221kPa 7:59
VENTING 7:59
AIR DRYING 8:50
CYCLE COMPLETE 9:10

Digital Signature #
F9B158EF2A48471A
