
Midmark M11 - v1.0.5
Total Cycles: 3280

Name: RITTER2
Sterilizer ID: V2313228
01/10/2024 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.2 F 0.0
2:00 75.4 F 0.0
4:00 89.3 F 0.1
6:00 110.7 F 1.1
8:00 144.1 F 3.8
10:00 171.1 F 6.7
12:00 186.6 F 10.4
14:00 204.3 F 7.1
16:00 215.5 F 6.4
18:00 226.0 F 10.1
20:00 232.4 F 9.6
22:00 238.3 F 11.3
24:00 236.1 F 9.5
26:00 249.2 F 15.7
28:00 260.2 F 21.9
30:00 270.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.7 F 30.0
2:00 271.5 F 29.3
3:00 272.7 F 30.1
4:00 271.8 F 29.3

Min 271.5 F 29.2
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.5 F 0.0
5:00 227.8 F 0.0
10:00 195.7 F 0.0
15:00 190.3 F 0.0
20:00 186.2 F 0.0
25:00 183.7 F 0.0
30:00 182.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:49
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:08:56
CYCLE COMPLETE

01/10/2024 10:22 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____