
Midmark M11 - v1.0.5
Total Cycles: 2494

Name: RITTER2
Sterilizer ID: V2313228
06/09/2023 11:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.7 F 0.0
2:00 145.7 F 2.2
4:00 152.8 F 5.3
6:00 180.9 F 9.6
8:00 205.9 F 7.0
10:00 219.7 F 8.6
12:00 232.3 F 12.3
14:00 235.3 F 10.7
16:00 234.1 F 8.7
18:00 247.9 F 15.2
20:00 259.7 F 21.7
22:00 270.2 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.6
1:00 272.6 F 30.3
2:00 271.5 F 29.6
3:00 272.0 F 30.3
4:00 272.4 F 30.1

Min 271.2 F 29.5
Max 272.7 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 218.3 F 0.0
10:00 225.3 F 0.0
15:00 218.1 F 0.0
20:00 218.2 F 0.0
25:00 218.3 F 0.0
30:00 217.5 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 22:54
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:01:15
CYCLE COMPLETE

06/09/2023 12:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____