
Midmark M11 - v1.0.0
Total Cycles: 3483

Name: RITTER 1
Sterilizer ID: V2178198
07/04/2023 03:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.5 F 0.0
2:00 179.9 F 3.7
4:00 196.4 F 8.2
6:00 221.7 F 11.0
8:00 226.4 F 8.4
10:00 238.6 F 12.1
12:00 240.1 F 11.1
14:00 238.2 F 9.6
16:00 252.2 F 16.4
18:00 264.0 F 23.3
20:00 271.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 28.9
2:00 271.4 F 28.2
3:00 272.6 F 29.1
4:00 272.1 F 28.6

Min 271.3 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 218.1 F 0.0
10:00 232.4 F 0.0
15:00 231.9 F 0.0
20:00 227.8 F 0.0
25:00 230.1 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:02
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

07/04/2023 04:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____