
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000211
8:05 22/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:32
PRESSURIZING 4:32
STERILIZING 5:59
136.5 C 233kPa 5:59
136.3 C 233kPa 6:29
136.4 C 232kPa 6:59
136.4 C 233kPa 7:29
136.2 C 231kPa 7:59
136.4 C 232kPa 8:29
136.3 C 233kPa 8:59
136.3 C 232kPa 9:29
136.3 C 232kPa 9:29
VENTING 9:29
AIR DRYING 10:39
CYCLE COMPLETE 12:08

Digital Signature #
BDDCE9DFEBEA58AF
