
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005874
7:52 30/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:48
STERILIZING 3:55
136.5 C 223kPa 3:55
136.2 C 222kPa 4:25
136.3 C 222kPa 4:55
136.4 C 223kPa 5:25
136.2 C 222kPa 5:55
136.1 C 217kPa 6:25
136.4 C 222kPa 6:55
136.2 C 220kPa 7:25
136.2 C 220kPa 7:25
VENTING 7:25
AIR DRYING 8:13
CYCLE COMPLETE 28:41

Digital Signature #
83BFD94125B94F87
