
Midmark M11 - v1.0.5
Total Cycles: 1205

Name: RITTER2
Sterilizer ID: V2313228
06/21/2022 09:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.1 F 0.0
2:00 156.1 F 2.3
4:00 166.2 F 5.7
6:00 197.3 F 10.4
8:00 220.5 F 6.1
10:00 225.2 F 10.9
12:00 231.4 F 7.6
14:00 242.1 F 12.9
16:00 243.0 F 12.2
18:00 254.9 F 18.6
20:00 266.4 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 271.9 F 29.7
2:00 272.2 F 29.3
3:00 271.3 F 29.3
4:00 273.1 F 30.4

Min 270.9 F 28.9
Max 273.2 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 230.9 F 0.0
10:00 231.5 F 0.0
15:00 228.1 F 0.0
20:00 226.5 F 0.0
25:00 225.1 F 0.0
30:00 207.4 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 21:34
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:17
CYCLE COMPLETE

06/21/2022 10:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____