
Midmark M11 - v1.0.5
Total Cycles: 3396

Name: RITTER2
Sterilizer ID: V2313228
02/21/2024 12:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.9 F 0.0
2:00 139.1 F 0.9
4:00 144.8 F 3.4
6:00 173.7 F 7.2
8:00 196.9 F 11.5
10:00 215.2 F 5.5
12:00 228.3 F 10.5
14:00 230.5 F 7.5
16:00 241.5 F 13.1
18:00 241.9 F 11.9
20:00 254.5 F 18.4
22:00 265.6 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.6 F 29.6
2:00 271.7 F 29.8
3:00 272.0 F 29.4
4:00 272.0 F 30.0

Min 271.1 F 29.1
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.2 F 0.0
5:00 210.6 F 0.0
10:00 212.1 F 0.0
15:00 214.9 F 0.0
20:00 210.4 F 0.0
25:00 204.4 F 0.0
30:00 200.9 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 23:42
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:02:15
CYCLE COMPLETE

02/21/2024 01:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____