
Midmark M11 - v1.0.0
Total Cycles: 4979

Name: RITTER 1
Sterilizer ID: V2178198
10/09/2024 08:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 72.6 F 0.0
4:00 82.1 F 0.0
6:00 111.8 F 0.1
8:00 142.1 F 0.6
10:00 171.0 F 2.8
12:00 188.5 F 5.6
14:00 205.0 F 9.1
16:00 217.2 F 12.8
18:00 222.1 F 5.9
20:00 231.7 F 10.1
22:00 234.4 F 8.9
24:00 241.5 F 11.5
26:00 237.6 F 9.4
28:00 250.7 F 15.6
30:00 261.5 F 21.7
32:00 271.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.9
2:00 271.8 F 28.2
3:00 272.1 F 28.7
4:00 272.4 F 28.8

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.4 F 0.0
5:00 206.2 F 0.0
10:00 195.4 F 0.0
15:00 189.8 F 0.0
20:00 222.5 F 0.0
25:00 221.7 F 0.0
30:00 222.1 F 0.0

Drying Complete

FILLING: 0:43
HEATING: 32:41
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:10:55
CYCLE COMPLETE

10/09/2024 09:23 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____