

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4033

Name: RITTER2  
Sterilizer ID: V2313228  
10/16/2024 01:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 122.6 F 0.0  
2:00 122.2 F 0.1  
4:00 129.4 F 1.6  
6:00 160.2 F 4.8  
8:00 183.1 F 8.4  
10:00 207.4 F 12.3  
12:00 216.8 F 6.8  
14:00 229.5 F 11.0  
16:00 229.3 F 7.3  
18:00 239.7 F 12.7  
20:00 239.6 F 11.1  
22:00 251.3 F 16.9  
24:00 261.5 F 23.0  
26:00 271.2 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.7  
1:00 272.6 F 30.2  
2:00 271.5 F 29.7  
3:00 272.5 F 30.3  
4:00 271.8 F 29.7

Min 271.4 F 29.6  
Max 272.7 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.2 F 0.0  
5:00 210.7 F 0.0  
10:00 216.8 F 0.0  
15:00 220.7 F 0.0  
20:00 224.1 F 0.0  
25:00 227.2 F 0.0  
30:00 229.2 F 0.0

Drying Complete

FILLING: 0:57  
HEATING: 26:41  
STERILIZING: 4:00  
VENTING: 3:53  
DRYING: 30:00  
TOTAL CYCLE: 01:05:30  
CYCLE COMPLETE

10/16/2024 02:52 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_