
Midmark M11 - v1.0.0
Total Cycles: 2003

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2022 01:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.8 F 0.0
2:00 160.4 F 3.2
4:00 182.6 F 7.3
6:00 209.9 F 12.8
8:00 220.0 F 6.3
10:00 232.5 F 11.1
12:00 231.2 F 7.4
14:00 244.5 F 13.5
16:00 242.8 F 11.9
18:00 255.7 F 18.6
20:00 267.0 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.5 F 29.2
2:00 271.3 F 28.5
3:00 272.5 F 29.3
4:00 271.6 F 28.6

Min 271.2 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 208.4 F 0.0
10:00 223.4 F 0.0
15:00 227.4 F 0.0
20:00 230.2 F 0.0
25:00 228.2 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

06/07/2022 02:46 PM
ACCEPTED: X
REJECTED:
Operator: U02