

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5314

Name: RITTER 1  
Sterilizer ID: V2178198  
02/04/2025 01:39 PM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	70.8 F	0.0
2:00	77.5 F	0.4
4:00	102.8 F	2.0
6:00	133.1 F	4.8
8:00	157.3 F	8.1
10:00	178.7 F	12.2
12:00	201.6 F	2.7
14:00	215.3 F	6.6
16:00	226.9 F	10.4
18:00	229.6 F	7.4
20:00	238.5 F	10.7
22:00	234.0 F	7.6
24:00	248.6 F	14.5
26:00	260.6 F	21.1

C103: HEATUP MODE  
STOP PRESSED  
CHAMBER TEMP 256F  
CHAMBER PRESS 18.4PSI  
\*\* ITEMS NOT STERILE \*\*

02/04/2025 02:07 PM