
Midmark M11 - v1.0.0
Total Cycles: 3808

Name: RITTER 1
Sterilizer ID: V2178198
09/25/2023 11:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.0 F 0.0
2:00 189.8 F 2.2
4:00 191.0 F 6.3
6:00 208.3 F 10.8
8:00 225.1 F 5.2
10:00 228.8 F 9.7
12:00 237.4 F 9.6
14:00 241.0 F 11.7
16:00 239.0 F 10.1
18:00 251.8 F 16.1
20:00 262.5 F 22.3
22:00 272.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.4 F 28.9
2:00 271.5 F 28.2
3:00 272.3 F 29.0
4:00 272.1 F 28.6

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 223.6 F 0.0
10:00 228.8 F 0.0
15:00 230.7 F 0.0
20:00 224.7 F 0.0
25:00 225.2 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 22:29
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:01:33
CYCLE COMPLETE

09/25/2023 12:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____