
Midmark M11 - v1.0.0
Total Cycles: 2102

Name: RITTER 1
Sterilizer ID: V2178198
07/07/2022 04:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.4 F 0.0
2:00 156.3 F 2.4
4:00 180.0 F 6.3
6:00 198.3 F 11.4
8:00 217.0 F 5.1
10:00 230.8 F 10.2
12:00 232.4 F 7.7
14:00 243.5 F 12.7
16:00 241.6 F 11.1
18:00 254.8 F 17.8
20:00 266.4 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.5 F 28.7
2:00 271.9 F 28.6
3:00 272.6 F 28.6
4:00 272.0 F 28.6

Min 271.6 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.2 F 0.0
5:00 201.1 F 0.0
10:00 196.5 F 0.0
15:00 192.4 F 0.0
20:00 188.6 F 0.0
25:00 185.5 F 0.0
30:00 184.0 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 21:33
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:00:06
CYCLE COMPLETE

07/07/2022 05:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____