
Midmark M11 - v1.0.5
Total Cycles: 3436

Name: RITTER2
Sterilizer ID: V2313228
03/06/2024 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.0 F 0.0
2:00 93.6 F 0.0
4:00 100.8 F 0.2
6:00 126.3 F 2.3
8:00 158.1 F 5.6
10:00 182.0 F 9.0
12:00 199.8 F 12.8
14:00 212.1 F 5.9
16:00 222.2 F 10.0
18:00 232.3 F 9.5
20:00 238.6 F 11.7
22:00 237.7 F 10.0
24:00 250.5 F 16.0
26:00 260.8 F 22.1
28:00 271.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.6
1:00 272.5 F 29.6
2:00 272.7 F 29.9
3:00 272.3 F 30.2
4:00 271.9 F 29.5

Min 271.8 F 29.4
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.8 F 0.0
5:00 216.9 F 0.0
10:00 201.7 F 0.0
15:00 190.6 F 0.0
20:00 188.0 F 0.0
25:00 184.9 F 0.0
30:00 182.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:34
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:07:59
CYCLE COMPLETE

03/06/2024 04:48 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____